Petrifilm™

Created by Authorized Personnel: Lola Johnson, 2019-10-07

Manufacture Date: 2019-10-01
Expiration Date: 2021-03-31

Product Manufacturing Certificate

Certificate of Analysis

Product: 3M™ Petrifilm™ Aerobic Count Plates 6400, 6403, 6406 or 6442
Batch: 335LKT
Stock Number: 70-2005-7212-4 or 70-2006-7784-0 or 70-2005-7215-7 or 70-2007-7071-0
ERP Number: 7100039310 or 7100048022 or 7100039374 or 7100047858

Organism Tested	Growth Specification	Result
Lactococcus lactis subsp. *cremoris* ATCC 19257	≥-3.0*	Pass
Escherichia coli ATCC 51813	≥-3.0*	Pass

*Expressed as the number of standard deviations away from the average count on standard agar medium.

ISO 11133:2014 Performance Testing

Organism Tested	Growth Specification	Result
Escherichia coli ATCC 25922 (WDCM 00013)	Productivity Ratio≥0.7	Pass
Bacillus subtilis subsp. *spizizenii* ATCC 6633 (WDCM 00003)	Productivity Ratio≥0.7	Pass
Escherichia coli ATCC 8739 (WDCM 00012)	Productivity Ratio≥0.7	Pass
Staphylococcus aureus ATCC 25923 (WDCM 00034)	Productivity Ratio≥0.7	Pass
Staphylococcus aureus ATCC 6538 (WDCM 00058)	Productivity Ratio≥0.7	Pass

This material complies with the 3M specifications for this product construction, and applicable criteria for routine quality control and microbiological performance of ISO 11133. 3M Brookings is certified to ISO 9001 through an independent agency.

3M Health Care
3M Center, Building 275-5W-05
St. Paul, MN 55144-1000 Phone: 1-800-328-1671

3M and Petrifilm are trademarks of 3M. Please recycle.
Printed in USA © 3M 2017. All rights reserved.

Version 8

AC
Aerobic Count Plate