



# Petrifilm™

Created by Authorized Personnel: Katarzyna Mamzer 2019-08-09

Manufacture Date: 2019-07-31

Expiration Date: 2021-01-28

## Product Manufacturing Certificate

## Certificate of Analysis

Product: 3M™ Petrifilm™ Coliform Count Plates 6410 or 6416

Batch: 418319212A

Stock Number: UU-0081-9644-4 or UU-0081-9648-5

ERP Number: 7100126846 or 7100126806

Organism Tested	Growth Specification	Result
<i>Escherichia coli</i> ATCC 51813	≥-3.0*, gas	Pass
<i>Enterobacter amnigenus</i> ATCC 51816	≥-3.0*, gas	Pass
<i>Klebsiella oxytoca</i> ATCC 51817	≥-3.0*, gas	Pass
<i>Enterococcus faecalis</i> ATCC 14506	No Growth	Pass

\*Expressed as the number of standard deviations away from the average count on standard agar media

### ISO 11133:2014 Performance Testing

Organism Tested	Growth Specification	Result
<i>Escherichia coli</i> ATCC 25922 (WDCM 00013)	Productivity Ratio≥0.5	Pass
<i>Escherichia coli</i> ATCC 8739 (WDCM 00012)	Productivity Ratio≥0.5	Pass
<i>Enterococcus faecalis</i> ATCC 29212 (WDCM 00087)	No Growth	Pass
<i>Enterococcus faecalis</i> ATCC 19433 (WDCM 00009)	No Growth	Pass
<i>Pseudomonas aeruginosa</i> ATCC 27853 (WDCM 00025)	Atypical Growth	Pass

This material complies with the 3M specifications for this product construction, and applicable criteria for routine quality control and microbiological performance of ISO 11133. 3M Wroclaw is certified to ISO 9001 through an independent agency.

3M Health Care  
3M Center, Building 275-5W-05  
St. Paul, MN 55144-1000 Phone: 1-800-328-1671

3M and Petrifilm are trademarks of 3M. Please recycle.  
Printed in USA © 3M 2017. All rights reserved.

Version 1



Coliform Count Plate