

Created by Authorized Personnel: Lola, 2019-03-29

Manufacture Date: 2019-03-20 Expiration Date: 2020-09-17

Product Manufacturing Certificate

Certificate of Analysis

Product: 3M[™] Petrifilm[™] Salmonella Express Plate 6536 or 6537

Batch: 3347LC

Stock Number: 70-2007-6998-5 or 70-2007-6999-3

ERP Number: 7100039574 or 7100039575

| Organism Tested | Minimum Growth | Result |
|---|-----------------|--------|
| Salmonella enterica subsp enterica Typhimurium ATCC 14028 | Typical Growth | Pass |
| Salmonella enterica subsp enterica Enteritidis ATCC 49223 | Typical Growth | Pass |
| Citrobacter freundii FSD 548 | Atypical Growth | Pass |
| Escherichia coli FSD 723 | Atypical Growth | Pass |

This material complies with the 3M specifications for this product construction, and applicable criteria for routine quality control and microbiological performance. 3M Brookings is certified to ISO 9001 through an independent agency.

3M Health Care 3M Center, Building 275-5W-01 St. Paul, MN 55144-1000 Phone: 1-800-328-1671

3M and Petrifilm are trademarks of 3M. Please recycle. Printed in USA @ 3M 2017. All rights reserved.

Version 2

